

Certificate of Analysis

Name: **RAMP® Troponin I** Batch Number: **M07318** Use-by: **2019-07-31**

Lot Release Specifications:

Criteria	Specification	Result
Barcode Detected	Yes	Yes
Error Messages	None observed	None observed
Accuracy	Mean Results for Accuracy Sample within +/-20% of assigned value	0.25 ng/mL
Precision	Result CV for Precision Sample must be $\leq 10\%$	8 %
Standard Curve Plot	$R^2 > 0.990$	0.999
Cardiac Controls	Within Range	Met Requirement

Conforms to product release specifications of Response Biomedical Corporation

Release Date: 31 – Jan – 2018

Signature: 
Quality Assurance

Date: 31 Jan 18.