

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2015-11
KIT DESCRIPTION:	ALT
CATALOGUE NO:	AL3801
BATCH NO:	382625
EXPIRY DATE:	2016-12

SAMPLE	TARGET	RANGE	RESULT	UNITS
CONTROL I / 1851CH	85.7	70.3 – 101.0	87.0	g/L
CONTROL III / 1856CH	157	129 – 185	156	g/L

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE