

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-11
KIT DESCRIPTION:	Total Bilirubin Reagent
CATALOGUE NO:	BR3859
BATCH NO:	466905
EXPIRY DATE:	2021-10-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II/ 1237UN	32.1	25.3 – 38.9	32.2	μmol/L
Level III/ 766UE	87.0	68.7 – 105.0	90.2	µmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | I www.randox.com