



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-01
KIT DESCRIPTION:	ALT IFCC
CATALOGUE NO:	AL3801
BATCH NO:	434241
EXPIRY DATE:	2019-03

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 892UN	38	30-46	37	U/L
LEVEL III / 766UE	144	115-173	139	U/L

APPROVED BY:

Edgar M. Snodgrass

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Micro, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 030-000.

4706-QC Rev4
27th September 2013