

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2015-09
KIT DESCRIPTION:	RANSOD
CATALOGUE NO:	SD125
BATCH NO:	358385
EXPIRY DATE:	2017-07

SAMPLE	TARGET	RANGE	RESULT	UNITS
326RD	348	285-411	379	U/mL
315RD	266	218-314	293	U/mL

**APPROVED BY:** 

**QUALITY CONTROL REPRESENTATIVE** 

vette Millar