

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-10
KIT DESCRIPTION:	GAMMA GT
CATALOGUE NO:	GT3874
BATCH NO:	396966
EXPIRY DATE:	2017-10

SAMPLE	TARGET	RANGE	RESULT	UNITS
HUMAN NORMAL SERUM 892UN	54	46-62	54	U/L
HUMAN ELEVATED SERUM 766UE	170	145-195	166	U/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com