



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-04
KIT DESCRIPTION:	LIPASE REAGENT
CATALOGUE NO:	LI188
BATCH NO:	411841
EXPIRY DATE:	2018-04

SAMPLE	TARGET	RANGE	RESULT	UNITS
892UN / LEVEL II	170	136 - 204	174	U/L
851UE / LEVEL III	430	345 - 515	459	U/L

APPROVED BY:

R. Ball

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAC Man. R J Fitzgerald W Fitzgerald CBE R F Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI6738, VAT Number GB 353 0904 00.

4706-QC Rev4
27th September 2013