



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-11
KIT DESCRIPTION:	AST
CATALOGUE NO:	AS8306
BATCH NO:	433787
EXPIRY DATE:	2019-02-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II / 892UN	42	34-50	40	U/L
Level III / 766UE	145	116-174	149	U/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRCR FRAC, R J Fitzgerald, W Fitzgerald CBE, R P Kelly BSc PCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI5738, VAT Number GB 353 030100.

4706-QC Rev4
27th September 2013