



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                      |         |
|----------------------|---------|
| DATE OF MANUFACTURE: | 2016-09 |
| KIT DESCRIPTION:     | FrCRP   |
| CATALOGUE NO:        | CP3847  |
| BATCH NO:            | 396436  |
| EXPIRY DATE:         | 2017-07 |

| SAMPLE                             | TARGET | RANGE     | RESULT | UNITS |
|------------------------------------|--------|-----------|--------|-------|
| SP CONTROL<br>LEVEL II/<br>454LPC  | 47.4   | 37.9-56.9 | 44.73  | Mg/L  |
| SP CONTROL<br>LEVEL III/<br>455LPC | 70.3   | 56.2-84.8 | 70.34  | Mg/L  |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRCR FRAC R J Fitzgerald M R Fitzgerald CBE R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI5738, VAT Number GB 353 0269 00

4706-QC Rev4  
27<sup>th</sup> September 2013