

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-06
KIT DESCRIPTION:	СК-МВ
CATALOGUE NO:	CK1296
BATCH NO:	402355
EXPIRY DATE:	2019-12-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
CK-MB Control / 3821CK	158	134-182	151	U/L
CK-MB Control / 4077CK	156	133-179	150	U/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE