DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: TOXOPLASMA IgG - CHORUS

Code 81040

Lot No. 951 **Expiry** 2014-04

REAGENTS

Label	Description	Lot.	Exp	No
D2271	TOXOPLASMA IgG	951	2014-04	6X6
B2272	CALIBRATORE TOXOPLASMA IgG	951	2014-04	1
C2271	CONTROL SERUM TOXOPLASMA IgG	951	2014-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	280	≥ 100
Control Serum (IU/mL)	51,0	30,0 - 77,8
Negative Control (IU/mL)	0,0	<5

CALIBRATOR

0	8	2	9	5	1	5	2	2	3	3	8	3	2	5	6	0	0

Label checked with barcode reader YES [x] Operator Signature .....



CONTROL SERUM

0	8	2	9	5	1	5	2	6	4	8	0	0	0	0	0	0	1

Label checked with barcode reader YES [x] Operator Signature .....

The test has been carried out in accordance with the package insert

TESTED BY: Laura Capitani

APPROVED RESPONSIBLE HEAD,Quality Control Reagents.....

DATE 31/07/2013

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.