

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-01
KIT DESCRIPTION:	CHOLESTEROL
CATALOGUE NO:	CH3810
BATCH NO:	434528
EXPIRY DATE:	2019-11-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II 892UN	4.39	3.82 – 4.96	4.29	MMOL/L
LEVEL III 766UE	6.68	5.81 – 7.55	6.60	MMOL/L

APPROVED BY:

**QUALITY CONTROL REPRESENTATIVE** 

'ener Danuelly