

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-04
KIT DESCRIPTION:	CRP CALIBRATOR SERIES
CATALOGUE NO:	CP2499
BATCH NO:	413359
EXPIRY DATE:	2018-09-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 401LPC	FRCRP	24.0	19.2-28.8	22.9	mg/L
LEVEL II / 402LPC	FRCRP	48.4	38.7-58.1	44.5	mg/L

APPROVED BY:

Lisa Ren

QUALITY CONTROL REPRESENTATIVE