RANDOX

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-08
KIT DESCRIPTION:	Full Range CRP
CATALOGUE NO:	CP3847
BATCH NO:	427258
EXPIRY DATE:	2019-05-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level I / 450LPC	23.2	18.6-27.8	21.5	mg/L
Level III / 452LPC	68.9	55.1-82.7	67.2	mg/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE