

Certificate of Analysis

Product Name: Coag Control Level 3 (Abnormal)-Orig:USA
Product Number: 100597
Lot Number: 652728

Expiration Date: 06/30/2017
Storage: 2 - 8°C

Specification	Specification Limit	Result
Appearance	Yellow to buff colored solid	Complies
Appearance Reconstituted	Yellow to straw colored, clear to hazy, free of particulates when viewed against dark background	Complies
Solubility	Less than 2 minutes	Complies
Vial-to-Vial Variability	≤ 4 %	3
Moisture	≤ 3 %	1
Thromboplastin D	26.0 - 34.0 S	31.0
Thromboplastin D	Confidence Interval*	26.3-35.6 s
Thromboplastin D	Value % PT	12 %
Thromboplastin D	Confidence Interval*	10-15 %
APTT-XL	65 - 79 S	72
APTT-XL	Confidence Interval*	61-83 s
Kontakt	Report	57 S
Kontakt	Confidence Interval*	48-65 s
Fibrinogen**	150 - 450 mg/dL	282
Fibrinogen**	Confidence Interval*	240-325 mg/dL

For in-vitro diagnostic use.

* Calculated from recovered values. Actual values recovered in the user laboratory will depend on reagent lot number, instrument and technique in use. Each laboratory should establish ranges using instrument, reagent and technique in place.

** Clauss fibrinogen

All individual donor units comprising this lot of Coagulation Control Plasma have been tested and found nonreactive for HBsAg and negative for antibodies to HIV and HCV. As with all blood products, this product should be handled as potentially infectious biological material.



Quality Representative

Date: 2015-04-13