DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: HERPES SIMPLEX 2 IgG RECOMBINANT - CHORUS

Code 81029

Lot No. 290 **Expiry** 2017-08

REAGENTS

Label	Description	Lot.		No
D8690	Herpes simplex 2 IgG Recombinant	290	2017-08	6X6
B8690	CALIBRATORE Herpes simplex 2 IgG Recombinant	290	2017-08	1
C8690	CONTROL SERUM Herpes simplex 2 IgG Recombinant	290	2017-08	1

CONTROLS

	Found results	Ex
Calibrator (mOD)	530	≥ 100
Control Serum (INDEX)	2,2	1,4-2,6
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

2	6	0	2	9	0	5	0	0	0	0	0	5	0	0	0	0	0

Label checked with barcode reader **YES** ☒ Operator Signature .....AB....



CONTROL SERUM

2	6	0	2	9	0	5	1	4	2	6	0	0	0	0	0	0	1

Label checked with barcode reader **YES** ☒ Operator Signature .....AB....

The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci

APPROVED RESPONSIBLE HEAD,Quality Control Reagents .....Acionelli.....

DATE 02/12/2016

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.