

RANDEX QUALITY CONTROL CERTIFICATE OF ANALYSIS

This is to certify that the following reagents were tested and approved by Randox Quality Control Department.

DATE OF MANUFACTURE: 2016-01
KIT DESCRIPTION: BILIARY ACIDS
CATALOGUE NO: BI2672
BATCH NO: 392772
EXPIRY DATE: 2016-12

SAMPLE	TARGET	RANGE	RESULT
LEVEL II / 482UN	0.84	0.71 – 0.96	0.80mmol/l
LEVEL III / 355UE	0.60	0.51 – 0.69	0.54mmol/l
LINEARITY:	2.00mmol/l		

APPROVED BY:



QUALITY CONTROL ANALYST