

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-06
KIT DESCRIPTION:	LIPASE REAGENT
CATALOGUE NO:	LI188
BATCH NO:	420040
EXPIRY DATE:	2019-05

SAMPLE	TARGET	RANGE	RESULT	UNITS
892UN / LEVEL II	170	136 - 204	179	U/L
766UE / LEVEL III	415	333 - 497	536	U/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com