DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: VARICELLA IgG - CHORUS

Code 81078

Lot No. 371 2018-09 2017-09

REAGENTS

Label	Description	Lot.	Exp	No
D2471	VARICELLA IgG	371	2018-09	6x6
B2472	CALIBRATORE VARICELLA IgG	371	2018-09	1
C2471	CONTROL SERUM VARICELLA IgG	371	2018-09	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	389	≥ 100
CONTROL SERUM (Index)	3,3	2,4-4,4
NEGATIVE CONTROL (Index)	0,0	< 0,7

CALIBRATOR

0	1	4	3	7	1	6	0	0	0	0	0	3	0	0	0	0	0

Label checked with barcode reader **YES** ☒ Operator Signature .....

CONTROL SERUM

0	1	4	3	7	1	6	2	4	4	4	0	0	0	0	0	0	1

Label checked with barcode reader **YES** ☒ Operator Signature .....

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD,Quality Control Reagents.....

DATE 13/11/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.