

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: Epstein-Barr VCA IgM II - CHORUS

Code	81054		
Lot No.	328	2018-07	2017-04

REAGENTS

Label	Description	Lot.	Exp	No
D8490	Epstein-Barr VCA IgM II	328	2018-07	6x6
B9490	CALIBRATOR Epstein-Barr VCA IgM II	328	2018-07	1
C9490	CONTROL SERUM Epstein-Barr VCA IgM II	328	2018-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	437	≥ 100
Control Serum (INDEX)	2,5	1,9-3,5
Negative Control (INDEX)	0,0	< 0,7

CALIBRATOR

0	3	0	3	2	8	7	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

 03032870000400000 Epstein-Barr VCA IgM II CALIBRATOR 0.175 mL IVD LOT 328 2018-07 2°C-8°C DIESSE Monteriggioni - Italy	
--	--

CONTROL SERUM

0	3	0	3	2	8	7	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

 030328719350000001 Epstein-Barr VCA IgM II CONTROL 0.425 mL IVD LOT 328 2018-07 2°C-8°C DIESSE Monteriggioni - Italy	
--	--

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED: RESPONSIBLE HEAD, Quality Control Reagents

DATE: 08/06/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.