

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| DATE OF MANUFACTURE: | 2017-07              |
|----------------------|----------------------|
| KIT DESCRIPTION:     | AMMONIA ENZYMATIC UV |
| CATALOGUE NO:        | AM1054               |
| BATCH NO:            | 418986               |
| EXPIRY DATE:         | 2019-05-28           |

| SAMPLE               | TARGET | RANGE   | RESULT | UNITS  |
|----------------------|--------|---------|--------|--------|
| LEVEL II /<br>639AM  | 142    | 114-170 | 136    | μmol/L |
| LEVEL III /<br>669AM | 288    | 230-346 | 297    | μmol/L |

APPROVED BY:

Viette Milla-

**QUALITY CONTROL REPRESENTATIVE**