

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-11-29
KIT DESCRIPTION:	COAGULATION CONTROL LEVEL 3
CATALOGUE NO:	CG5023
BATCH NO:	259CG
EXPIRY DATE:	2019-11-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
259CG	PROTHROMBIN TIME	30.2	N/A	26.9	SECS
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com