



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-04
KIT DESCRIPTION:	CHOLESTEROL LDL
CATALOGUE NO:	CH3841
BATCH NO:	449358
EXPIRY DATE:	2019-08-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I 2530CH	2.48	2.11 – 2.85	2.36	MMOL/L
LEVEL II 2532CH	3.74	3.18 – 4.30	3.54	MMOL/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Director: Dr S P Fitzgerald BSc PhD DSc (Hons) FRCR MRCP FRCR W Fitzgerald CBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI6728, VAT Number GB 353 0304 00.

4706-QC Rev4
27th September 2013