



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                      |                                        |
|----------------------|----------------------------------------|
| DATE OF MANUFACTURE: | 2018-01                                |
| KIT DESCRIPTION:     | CREATININE WITHOUT<br>DEPROTEINIZATION |
| CATALOGUE NO:        | CR510                                  |
| BATCH NO:            | 437397                                 |
| EXPIRY DATE:         | 2020-11                                |

| SAMPLE               | TARGET | RANGE       | RESULT | UNITS  |
|----------------------|--------|-------------|--------|--------|
| LEVEL II /<br>892UN  | 122.0  | 97.6-146.0  | 125.6  | mmol/L |
| LEVEL III /<br>766UE | 357.0  | 285.0-429.0 | 327.5  | mmol/L |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

**Randox Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

4706-QC Rev4  
27<sup>th</sup> September 2013