

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: September 2017

KIT DESCRIPTION: IA SPECIALITY 1 LEVEL 2

CATALOGUE NO: IAS3114

LOT NUMBERS: 1710EC

EXPIRY DATE: 2019-04-28

SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

| ANALYTE | LOT NUMBER | RANGE | TARGET | RESULT | PRECISION (if applicable) |
|-----------------------|------------|-------------|--------|--------|---------------------------|
| Anti TPO (kU/l) | 1709EC | 68.5-114 | 91.3 | 109.0 | N/A |
| | | | | 102.0 | |
| | | | | | |
| C-Peptide (nmol/l) | 1709EC | 1.020-1.700 | 1.360 | 1.281 | |
| | | | | 1.258 | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | _ | | | | |
| | | | | | |

DATE: 20 Sep 2017

RESULT (PASS/FAIL): Pass

QC APPROVED BY:

adèle Martin.