



CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: October 2017
KIT DESCRIPTION: IA Premium Plus Tri-Level
CATALOGUE NO: IA3112
LOT NUMBERS: 432364 (1714EC, 1716EC, 1719EC)
EXPIRY DATE: 2020-04-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Cortisol (nmol/l)	1714EC	161-268	214	213	N/A
				211	
FSH (mU/ml)	1714EC	5.30-7.96	6.63	7.0	N/A
				7.2	
LH (mU/ml)	1714EC	4.46-6.68	5.57	5.6	N/A
				5.6	
Cortisol (nmol/l)	1716EC	442-736	589	596	N/A
				588	
FSH (mU/ml)	1716EC	21.1-31.7	26.4	25.0	N/A
				26.4	
LH (mU/ml)	1716EC	32.4-48.6	40.5	36.1	N/A
				38.8	
Cortisol (nmol/l)	1719EC	622-1036	829	963	N/A
				858	
FSH (mU/ml)	1719EC	49.0-73.4	61.2	55.0	N/A
				58.8	
LH (mU/ml)	1719EC	54.9-82.3	68.6	63.5	N/A
				58.5	

QC TESTED BY: Ilona Palmer
DATE: 13 Oct 2017
RESULT (PASS/FAIL): Pass

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016



QC APPROVED BY:

Adele Martin.

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016