



Certificate of Analysis: Lyophilized Microorganism Specification and Performance Upon Release

Specifications Microorganism Name: Staphylococcus saprophyticus Catalog Number: 0134 Lot Number: 134-29 Reference Number: ATCC® BAA-750™* Purity: Pure Passage from Reference: 3	Expiration Date: 2019/1/31 Release Information: Quality Control Technologist: Christine Condon Release Date: 2017/3/7
---	--

Performance	
Macroscopic Features: Medium, circular, convex, entire edge, white, glistening Microscopic Features: Gram positive cocci singly, in pairs and tetrads	Medium: SBAP Method: Gram Stain (1)
ID System: Vitek GP (1) See attached ID System results document.	Other Features/ Challenges: Results (1) Catalase (3% Hydrogen Peroxide): positive (1) Coagulase (rabbit plasma-tube): negative  Amanda Kuperus Quality Control Manager AUTHORIZED SIGNATURE

Disclaimer: The last digit(s) of the lot number appearing on the product label and packing slip are merely a packaging event number. The lot number displayed on this certificate is the actual base lot number.

Note for Vitek®: Although the Vitek® panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods.

⚠ Refer to the enclosed product insert for instructions, intended use and hazard/safety information.

Individual products are traceable to a recognized culture collection.



(*) The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative word mark and the ATCC catalog marks are trademarks of ATCC. Microbiologics, Inc. is licensed to use these trademarks and to sell products derived from ATCC® cultures.



(1) These tests are accredited to ISO/IEC 17025:2005.

TESTING CERT #2655.01

bioMerieux Customer: 1012555
System #: 521711

Laboratory Report

Printed Mar 2, 2017 07:59 CST
Printed by: cmc

Isolate Group: 134 29-1

Bench: CC

Card Type: GP Testing Instrument: 000017919CE3 (15220)

Bionumber: 020000014670231

Organism Quantity:

Comments:	
	GP QC OK

Identification Information	Card: GP	Lot Number: 242399410	Expires: Nov 20, 2017 12:00 CST
	Completed: Feb 28, 2017 21:36 CST	Status: Final	Analysis Time: 5.00 hours
Selected Organism	99% Probability Staphylococcus saprophyticus Bionumber: 020000014670231 Confidence: Excellent identification		
SRF Organism			
Analysis Organisms and Tests to Separate:			
Analysis Messages:			
Contraindicating Typical Biopattern(s)			

Biochemical Details																	
2	AMY	-	4	PIPLC	-	5	dXYL	-	8	ADH1	-	9	BGAL	+	11	AGLU	-
13	APPA	-	14	CDEX	-	15	AspA	-	16	BGAR	-	17	AMAN	-	19	PHOS	-
20	LeuA	-	23	ProA	-	24	BGURr	-	25	AGAL	-	26	PyrA	-	27	BGUR	-
28	AlaA	-	29	TyrA	-	30	dSOR	-	31	URE	+	32	POLYB	-	37	dGAL	-
38	dRIB	-	39	ILATk	-	42	LAC	+	44	NAG	-	45	dMAL	+	46	BACI	+
47	NOVO	+	50	NC6.5	+	52	dMAN	+	53	dMNE	-	54	MBdG	-	56	PUL	-
57	dRAF	-	58	O129R	+	59	SAL	-	60	SAC	+	62	dTRE	+	63	ADH2s	-
64	OPTO	+															

Installed VITEK 2 Systems Version: 07.01

MIC Interpretation Guideline:

AES Parameter Set Name:

Therapeutic Interpretation Guideline:

AES Parameter Last Modified: