



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-06
KIT DESCRIPTION:	AMYLASE REAGENT
CATALOGUE NO:	AY3805
BATCH NO:	417848
EXPIRY DATE:	2018-11

SAMPLE	TARGET	RANGE	RESULT	UNITS
892UN / LEVEL II	100	85 - 115	90	U/L
766UE / LEVEL III	333	283 - 383	320	U/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR FRCR
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number No15738, VAT Number GB 353 0304 00.

4706-QC Rev4
27th September 2013