



CERTIFICATE OF CONFORMITY

Certificate for Lot Number: 50608

Product Code: TS/5-18

Product: Probact Amies charcoal (M40-A2) medium in tube.
Includes polystyrene shaft viscose tip swab in peel pouch (Pack Size 100)

Date Of Manufacture: 07/2016 Expiry: 2018-07 Shelf Life: 24 months

This is to certify that the product detailed on this certificate conforms to Technical Service Consultants Ltd procedures and specifications and has performed the tests detailed below within documented parameters.

Sterilisation Reference, Isotron PLC Number

S11666078-1-1 - 4155596 : Dose 26.5 - 39.7 kGy

For Irradiated products batch release is controlled by the verification of the radiation dose throughout the load.

Autoclave Reference, Time Manufactured

For autoclaved products, either a temperature datalogger or visual sterilisation indicator are used to confirm that sterility was achieved.

Physical and Microbiological Assurance

Tests Performed	Results Obtained
Batch Release (Sterility)	Pass
Validation of transport media according to M40-A2 including	Pass
Recovery of Indicator organism <i>P. aeruginosa</i> ATCC BAA-427	Pass
Recovery of Indicator organism <i>N. gonorrhoeae</i> ATCC 43069	Pass
Recovery of Indicator organism <i>B. fragilis</i> ATCC 25285	Pass

Overall Result: Pass

Certificate Issued by: Joanne Walsh Certificate authorised on 02.08.2016

Electronically generated on 24th October 2016 - valid without signature

Sent to Miguel Angel Rojas Cajamarca (compras@labcarecolombia.com)