

RANDOX

CERTIFICATE OF ANALYSIS


This is to certify that the following kits were tested by
Randox Testing Services and approved by the QC Department.

DATE OF MANUFACTURE: March 2014
KIT DESCRIPTION: Immunoassay Speciality I – Level 3
CATALOGUE NO: IAS3115
LOT NUMBERS: 1324EC
EXPIRY DATE: 2015-10
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Anti TPO (kIU/l)	1324EC	114 – 190	152	170	
				157	
Anti TG (kIU/l)	1324EC	152 – 254	203	198	
				187	

QC TESTED BY:



Janet Lightbody

DATE: 31-Mar-14

RESULT (PASS/FAIL): PASS

QC APPROVED BY:



Randox Laboratories Ltd. Ardmore, Diamond Road, Crumlin, Co. Antrim, United Kingdom, BT29 4QY
Tel : +44 2894 422413 Fax : +44 2894 452912

FORM NO. 6521-QC
REVISION (2)
22 OCTOBER 2008