

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI) - Italia - Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel +39-02-4859121 r.a Fax +39-02-48008530

# QUALITY CONTROL CERTIFICATE

KIT NAME: VARICELLA IgG - CHORUS

Code 81078

Lot No. 952 Expiry 2014-07

## REAGENTS

| Label | Description                 | Lot. | Exp     | No  |
|-------|-----------------------------|------|---------|-----|
| D2471 | VARICELLA IgG               | 952  | 2014-07 | 6x6 |
| B2472 | CALIBRATORE VARICELLA IgG   | 952  | 2014-07 | 1   |
| C2471 | CONTROL SERUM VARICELLA IgG | 952  | 2014-07 | 1   |

## CONTROLS

|                          | Found results | Expected results |
|--------------------------|---------------|------------------|
| CALIBRATOR (mOD)         | 320           | $\geq 100$       |
| CONTROL SERUM (Index)    | 2,8           | 1,7-3,1          |
| NEGATIVE CONTROL (Index) | 0,0           | $<0,7$           |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 1 | 4 | 9 | 5 | 2 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 4 | 2 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader

YES



Operator Signature

*JS*



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 1 | 4 | 9 | 5 | 2 | 6 | 1 | 7 | 3 | 1 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader

YES



Operator Signature

*JS*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

*S. Benocci*

DATE 29/07/2013

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.