



Certificate of Analysis: Lyophilized Microorganism Specification and Performance Upon Release

Specifications Microorganism Name: Escherichia coli Catalog Number: 0335 Lot Number: 335-348** Reference Number: ATCC® 25922™* Purity: Pure Passage from Reference: 3	Expiration Date: 2020/2/29 Release Information: Quality Control Technologist: Christine Condon Release Date: 2018/3/2
--	--

Performance	
Macroscopic Features: 2 colony types, both are gray & beta hemolytic; one is circular to irregular, convex, slightly erose edge & smooth; other is larger, irregular, low convex, erose edge & rough	Medium: SBAP
Microscopic Features: Gram negative straight rod	Method: Gram Stain (1)
ID System: MALDI-TOF (1) See attached ID System results document.	Other Features/ Challenges: Results (1) Oxidase (Kovacs): negative Beta-glucuronidase (E. coli Broth w/MUG): positive (1) Ampicillin (10 mcg - Disk Susceptibility): 16 - 22 mm (1) Gentamicin (10 mcg - Disk Susceptibility): 19 - 26 mm (1) SXT (1.25/23.75 mcg - Disk Susceptibility): 23 - 29 mm  Amanda Kuperus Quality Control Manager AUTHORIZED SIGNATURE

**Disclaimer: The last digit(s) of the lot number appearing on the product label and packing slip are merely a packaging event number. The lot number displayed on this certificate is the actual base lot number.

Note for Vitek®: Although the Vitek® panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods.

⚠ Refer to the enclosed product insert for instructions, intended use and hazard/safety information.

Individual products are traceable to a recognized culture collection.



(*) The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative word mark and the ATCC catalog marks are trademarks of ATCC. Microbiologics, Inc. is licensed to use these trademarks and to sell products derived from ATCC® cultures.

(1) These tests are accredited to ISO/IEC 17025:2005.

