



Certificate of Analysis: Lyophilized Microorganism Specification and Performance Upon Release

Specifications Microorganism Name: Klebsiella oxytoca Catalog Number: 0757 Lot Number: 757-58 Reference Number: ATCC® 700324™* Purity: Pure Passage from Reference: 3	Expiration Date: 2019/6/30 Release Information: Quality Control Technologist: Mary L Bowman Release Date: 2017/7/20
--	--

Performance	
Macroscopic Features: Large, circular, convex, entire edge, gray, glistening, opaque. Microscopic Features: Short, plump, gram negative, straight rod.	Medium: SBAP Method: Gram Stain (1)
ID System: Vitek GN (1) See attached ID System results document.	Other Features/ Challenges: Results (1) Oxidase (Kovacs): negative  Amanda Kuperus Quality Control Manager AUTHORIZED SIGNATURE

Disclaimer: The last digit(s) of the lot number appearing on the product label and packing slip are merely a packaging event number. The lot number displayed on this certificate is the actual base lot number.

Note for Vitek®: Although the Vitek® panel uses many conventional tests, the unique environment of the card, combined with the short incubation period, may produce results that differ from published results obtained by other methods.

⚠ Refer to the enclosed product insert for instructions, intended use and hazard/safety information.

Individual products are traceable to a recognized culture collection.



(*) The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative word mark and the ATCC catalog marks are trademarks of ATCC. Microbiologics, Inc. is licensed to use these trademarks and to sell products derived from ATCC® cultures.



(1) These tests are accredited to ISO/IEC 17025:2005.

TESTING CERT #2655.01

bioMerieux Customer: 1012555
System #: 521711

Laboratory Report

Printed Jul 13, 2017 09:09 CDT
Printed by: mlb

Isolate Group: 757 58 1-1

Bench: MB

Card Type: GN Testing Instrument: 000017919CE3 (15220)

Bionumber: 6705514276064001

Organism Quantity:

Comments:	

Identification Information	Card:	GN	Lot Number:	241399320	Expires:	Nov 19, 2017 12:00 CST
	Completed:	Jul 12, 2017 11:50 CDT	Status:	Final	Analysis Time:	2.00 hours
Selected Organism	99% Probability Klebsiella oxytoca					
	Bionumber: 6705514276064001				Confidence:	Excellent identification
SRF Organism						
Analysis Organisms and Tests to Separate:						
Analysis Messages:						
Contraindicating Typical Biopattern(s)						

Biochemical Details																	
2	APPA	-	3	ADO	+	4	PyrA	+	5	IARL	+	7	dCEL	+	9	BGAL	+
10	H2S	-	11	BNAG	-	12	AGLTp	-	13	dGLU	+	14	GGT	-	15	OFF	+
17	BGLU	+	18	dMAL	-	19	dMAN	+	20	dMNE	+	21	BXYL	-	22	BAlap	-
23	ProA	-	26	LIP	-	27	PLE	+	29	TyrA	-	31	URE	+	32	dSOR	-
33	SAC	+	34	dTAG	+	35	dTRE	+	36	CIT	-	37	MNT	+	39	5KG	+
40	ILATk	-	41	AGLU	-	42	SUCT	-	43	NAGA	-	44	AGAL	+	45	PHOS	+
46	GlyA	-	47	ODC	-	48	LDC	+	53	IHISa	-	56	CMT	-	57	BGUR	-
58	O129R	-	59	GGAA	-	61	IMLTa	-	62	ELLM	+	64	ILATa	-			

Installed VITEK 2 Systems Version: 07.01

MIC Interpretation Guideline:

AES Parameter Set Name:

Therapeutic Interpretation Guideline:

AES Parameter Last Modified: