

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: TOXOPLASMA IgG - CHORUS

Code 81040

Lot No. 372



2018-09



2017-09

REAGENTS

Label	Description	Lot.	Exp	No
D2271	TOXOPLASMA IgG	372	2018-09	6X6
B2272	CALIBRATORE TOXOPLASMA IgG	372	2018-09	1
C2271	CONTROL SERUM TOXOPLASMA IgG	372	2018-09	1

CONTROLS

	Found results	Expected result
--	---------------	-----------------

Calibrator (mOD)	642	≥ 100
Control Serum (IU/mL)	20,9	16,4-29,5
Negative Control (IU/mL)	0,0	< 5

CALIBRATOR


0	8	2	3	7	2	6	2	2	4	5	7	4	3	9	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *AF*



CONTROL SERUM


0	8	2	3	7	2	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *AF*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *AF*

DATE 06/12/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.