

DIESSE DIAGNOSTICA SENESE SpA
Via delle Rose 10, 53035 Monteriggioni (SI), Italy
Tel. ++39-0577-587111 fax ++39-0577-318690

Sede Legale e Amministrazione
Via A. Solari 19 20144 Milano Italy
Tel. ++39-02-4859121 r.a. fax ++39-02-48008530

CERTIFICATE OF ANALYSIS

<u>CODE</u>	<u>PRODUCT NAME</u>	<u>LOT</u>	<u>EXPIRY DATE</u>	<u>PREP DATE</u>
10439	SediCheck Cube 30 4 x 3 ml	014-A	2018-05	2017-05

VES-MATIC CUBE 30

Level I (NORMAL Level) lot 320	Test conditions	VES 1 st hour	Range of acceptability
VES-MATIC CUBE 30	Temperature OFF	4	1-15

Level II (ABNORMAL Level) lot 320	Test conditions	VES 1 st hour	Range of acceptability
VES-MATIC CUBE 30	Temperature OFF	70	55-98

MINI-CUBE

Level I (NORMAL Level) lot 320	Test conditions	VES 1 st hour	Range of acceptability
MINI- CUBE	Temperature ON	3	1-15

Level II (ABNORMAL Level) lot 320	Test conditions	VES 1 st hour	Range of acceptability
MINI- CUBE	Temperature ON	35	22-62

Date: 13/07/2017


Signature (Head Quality Control)