



DIAGNOSTIC REAGENTS

SeraQuest
Certificate of Analysis

Product: Anti-SSB IgG
Catalog No.: 01-260
Lot No.: A16018
Exp. Date: 2019-01-31

Reagent Blank Abs. 0.072
Calibrator 1 Abs. 1.472
Calibrator 2 Abs. 0.935

Microwell Plate:	Lot No. <u>A16118</u>	Exp. Date <u>Jan/19</u>	Accuracy <u>Passed</u> (QC Panel)
Conjugate:	Lot No. <u>A16218</u>	Exp. Date <u>Jan/19</u>	
Negative Control:	Lot No. <u>A16318</u>	Exp. Date <u>Jan/19</u>	Index <u>0.0</u>
Calibrator 1:	Lot No. <u>A16418</u>	Exp. Date <u>Jan/19</u>	Index <u>2.5</u>
Calibrator 2:	Lot No. <u>A16618</u>	Exp. Date <u>Jan/19</u>	Index <u>1.2</u>
Positive Control:	Lot No. <u>A16518</u>	Exp. Date <u>Jan/19</u>	Index <u>1.1 – 3.4</u>

Human source material used in preparation of this product have been tested by FDA licensed methods and found to be non-reactive for HBsAg, Anti-HCV, Anti-HIV 1 and Anti-HIV 2. No known test method can offer assurance that Hepatitis B Virus, Hepatitis C Virus, Human Immunodeficiency Virus or other infectious agents are absent. All human blood-based products should be handled in accordance with good laboratory practices and current acceptable guidelines for the prevention of blood-based viral infections.

Prepared By: [Signature]

Date: 02/19/2018

Certified By: [Signature]

Date: 02/19/2018

Reviewed By: [Signature]

Date: 02/19/2018

QUEST INTERNATIONAL, INC.
8127 N. W. 29th Street Doral, Florida 33122
Tel: 305-592-6991 Fax: 305-592-6834