

# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

This is to certify that the following reagents were tested and approved by Randox Quality Control Department.

|                                       |
|---------------------------------------|
| DATE OF MANUFACTURE: 2015-12          |
| KIT DESCRIPTION: ALKALINE PHOSPHATASE |
| CATALOGUE NO: AP3882                  |
| BATCH NO: 400252                      |
| EXPIRY DATE: 2017-01                  |

| SAMPLE            | TARGET    | RANGE     | RESULT      |
|-------------------|-----------|-----------|-------------|
| LEVELII<br>482UN  | 0.32      | 0.27-0.37 | 0.290mmol/l |
| LEVELIII<br>355UE | 1.51      | 1.28-1.74 | 1.35mmol/l  |
| LINEARITY:        | 3.2mmol/l |           |             |

APPROVED BY:

*U. Pandiarani*

QUALITY CONTROL ANALYST