

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: May 2018
KIT DESCRIPTION: IA SPECIALITY II LEV 3
CATALOGUE NO: IAS3119
LOT NUMBERS: 1777EC
EXPIRY DATE: 2019-12-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (ng/l)	1777EC	225 - 375	300	303	N/A
				309	
Gastrin (pmol/l)	1777EC	148 - 246	197	202	N/A
				187	

QC TESTED BY: Nicola McNamara
DATE: 02-May-18
RESULT (PASS/FAIL): Pass
QC APPROVED BY:

Adele Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15736. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016