

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-04
KIT DESCRIPTION:	HbA1c Reagent
CATALOGUE NO:	HA3830
BATCH NO:	453164
EXPIRY DATE:	2019-01-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level I/ 1848HA	5.75	4.60 – 6.90	5.71	%HbA1c
Level II/ 1901HA	12.20	9.76 – 14.60	12.08	%HbA1c

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com