DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: SS-B - CHORUS

Code 86024

Lot No. 003 **Expiry** 2015-01

REAGENTS

Label	Description	Lot.	Exp	No
D3201	SS-B	003	2015-01	6X6
B3201	CALIBRATORE SS-B	003	2015-01	1
C3201	CONTROL SERUM SS-B	003	2015-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	270	≥ 100
Control Serum (AU/mL)	44,2	28,3 - 60,0

CALIBRATOR

1	9	4	0	0	3	6	4	1	9	3	3	4	0	5	9	0	0

Label checked with barcode reader **YES** [X] Operator Signature .....



CONTROL SERUM

1	9	4	0	0	3	6	1	8	3	3	0	0	0	0	0	0	1

Label checked with barcode reader **YES** [X] Operator Signature .....



The test has been carried out in accordance with the package insert

TESTED BY: Andrea Bellavista

APPROVED RESPONSIBLE HEAD,Quality Control Reagents.....

DATE: 27/01/2014

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.