

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-05-28
KIT DESCRIPTION:	Direct HDL/LDL Cholesterol Calibrator
CATALOGUE NO:	CH2673
BATCH NO:	2437CH
EXPIRY DATE:	2019-05-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Calibrator / 2437CH	HDL Cholesterol	1.84	N/A	1.82	mmol/L
Calibrator / 2437CH	LDL Cholesterol	3.44	N/A	3.35	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com