

RANDEX QUALITY CONTROL CERTIFICATE OF ANALYSIS

This is to certify that the following reagents were tested and approved by Randox Quality Control Department.

DATE OF MANUFACTURE: 2015-06
KIT DESCRIPTION: AST
CATALOGUE NO: AS3804
BATCH NO: 394412
EXPIRY DATE: 2017-04

SAMPLE	TARGET	RANGE	RESULT
LEVEL II / 482UN	0.84	0.71 – 0.96	0.80mmol/l
LEVEL III / 355UE	0.60	0.51 – 0.69	0.54mmol/l
LINEARITY:	2.00mmol/l		

APPROVED BY:



QUALITY CONTROL ANALYST