

Certificate of Analysis

Name: **RAMP[®]
Procalcitonin** Batch Number: **M07363** Use-by: **2019-09-30**

Lot Release Specifications:

Criteria	Specification	Result
Barcode Detected	Yes	Yes
Error Messages	None observed	None
Standard Curve Plot	$R^2 > 0.990$	1.000
Precision	Result CV for Precision Sample must be $\leq 10\%$	10 %
PCT Controls	Within Range	Yes

Conforms to product release specifications of Response Biomedical Corporation

Release Date: 09 – Mar – 2018

Signature: *Theresa Kelly*
Quality Assurance

Date: 09 Mar 18