

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-11
KIT DESCRIPTION:	Urinary Protein
CATALOGUE NO:	UP1570
BATCH NO:	398035
EXPIRY DATE:	2017-11

SAMPLE	TARGET	RANGE	RESULT	UNITS
Control 2 / 773UC	0.120	0.096 - 0.144	0.140	g/L
Control 3/ 778UC	0.242	0.194 – 0.290	0.260	g/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE