

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: September 2017
KIT DESCRIPTION: IA SPECIALITY 1 LEVEL 1
CATALOGUE NO: IAS3113
LOT NUMBERS: 1709EC
EXPIRY DATE: 2019-04-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Anti TPO (kU/l)	1709EC	27.9-46.5	37.2	41.1	N/A
				42.4	
C-Peptide (nmol/l)	1709EC	0.318-0.530	0.424	0.437	
				0.430	

QC TESTED BY: Davina Graham
DATE: 20 Sep 2017
RESULT (PASS/FAIL): Pass
QC APPROVED BY: *Adele Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, P P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15736, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016