

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-05
KIT DESCRIPTION:	MICROALBUMIN
CATALOGUE NO:	MA8056
BATCH NO:	387328
EXPIRY DATE:	2018-04

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 891MA	31.4	25.1-37.7	31.8	mg/l
LEVEL II / 892MA	154.7	123.8-185.6	153.6	mg/l

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Atom there