

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: snRNP-C - CHORUS

Code 86092	
Lot No. 960	Expiry 2014-07

REAGENTS

Label	Description	Lot.	Exp	No
D7441	snRNP-C	960	2014-08	6X6
B7441	CALIBRATORE snRNP-C	960	2014-07	1
C7441	CONTROL SERUM snRNP-C	960	2014-07	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	220	≥ 100
Control Serum (AU/mL)	34,4	22,5 - 46,2

CALIBRATOR

1	9	1	9	6	0	6	2	2	0	3	3	4	8	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

Be



CONTROL SERUM

1	9	1	9	6	0	6	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

Be



The test has been carried out in accordance with the package insert

TESTED BY: Andrea Bellavista

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

M. Bontel

DATE: 06/08/2013

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.