



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                      |               |
|----------------------|---------------|
| DATE OF MANUFACTURE: | 2018-10       |
| KIT DESCRIPTION:     | TRIGLYCERIDES |
| CATALOGUE NO:        | TR3823        |
| BATCH NO:            | 461990        |
| EXPIRY DATE:         | 2021-08-28    |

| SAMPLE             | TARGET | RANGE       | RESULT | UNITS  |
|--------------------|--------|-------------|--------|--------|
| LEVEL II<br>1237UN | 1.15   | 0.97 – 1.33 | 1.08   | mmol/l |
| LEVEL III<br>766UE | 3.00   | 2.52 – 3.48 | 2.79   | mmol/l |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MAn. R J Fitzgerald. W Fitzgerald CBE. R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number GB 353 0304 00.

4706-QC Rev4  
27<sup>th</sup> September 2013