

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CYTOMEGALOVIRUS IgG -CHORUS

Code	81010		
Lot No.	317	Expiry	2018-03

REAGENTS

Label	Description	Lot.	Exp	No
D2161	Cytomegalovirus IgG	317	2018-03	6X6
B2162	CALIBRATORE Cytomegalovirus IgG	317	2018-03	1
C2161	CONTROL SERUM Cytomegalovirus IgG	317	2018-03	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	300	≥ 100
Control Serum (IU/mL)	7,4	5,2-9,0
Negative Control (IU/mL)	0,0	< 0,5

CALIBRATOR

0	0	1	3	1	7	6	2	2	2	4	7	3	1	5	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

CONTROL SERUM


0	0	1	3	1	7	6	2	3	4	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.....

Aliaudi

DATE

20/03/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.