



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-06
KIT DESCRIPTION:	LDH P-L
CATALOGUE NO:	LD8322
BATCH NO:	451444
EXPIRY DATE:	2019-08-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II 1237UN	424	361 – 487	425	U/L
LEVEL III 766UE	812	690 - 934	786	U/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Mar. R J Fitzgerald W Fitzgerald CBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.